

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012025\
 Data File : VV038460.D
 Acq On : 20 Jan 2025 13:39
 Operator : SY/MD
 Sample : Q1125-12
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B10

Quant Time: Jan 21 00:52:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 21 00:50:24 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	191380	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	206043	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	81213	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	80059	5.073	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	= 101.400%	
7) Chloroethane-d5	1.529	69	71077	5.363	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	= 107.200%	
11) 1,1-Dichloroethene-d2	2.056	65	35998	5.062	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	= 101.200%	
20) 2-Butanone-d5	3.789	46	161652	48.699	ug/L	-0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	= 97.400%	
24) Chloroform-d	4.239	84	155860	5.269	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 105.400%	
26) 1,2-Dichloroethane-d4	4.937	65	75261	5.552	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 111.000%	
32) Benzene-d6	4.953	84	291526	5.137	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	= 102.800%	
36) 1,2-Dichloropropane-d6	5.982	67	93458	5.418	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	= 108.400%	
41) Toluene-d8	7.236	98	229393	4.500	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 90.000%	
43) trans-1,3-Dichloroprop...	7.558	79	28996	4.596	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	= 92.000%	
46) 2-Hexanone-d5	8.021	63	103208	41.460	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	= 82.920%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	69636	5.092	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	= 101.800%	
66) 1,2-Dichlorobenzene-d4	11.554	152	79024	5.749	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	= 115.000%	
Target Compounds						Qvalue
16) Methylene chloride	2.445	84	3075	0.175	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012025\
Data File : VV038460.D
Acq On : 20 Jan 2025 13:39
Operator : SY/MD
Sample : Q1125-12
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B10

Quant Time: Jan 21 00:52:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 21 00:50:24 2025
Response via : Initial Calibration

