

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012221\
 Data File : VV020137.D
 Acq On : 22 Jan 2021 15:51
 Operator : SY/MD
 Sample : M1181-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 F1E39

Quant Time: Jan 23 00:15:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jan 23 00:14:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	377145	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.858	117	418962	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	171221	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.309	65	134423	48.06	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	96.12%		
7) Chloroethane-d5	1.569	69	112947	48.24	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	96.48%		
11) 1,1-Dichloroethene-d2	2.109	63	207491	37.14	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	74.28%		
21) 2-Butanone-d5	3.891	46	173448	90.41	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	90.41%		
24) Chloroform-d	4.354	84	241152	46.02	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.04%		
26) 1,2-Dichloroethane-d4	5.035	65	172287	47.44	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.88%		
32) Benzene-d6	5.055	84	452818	39.25	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	78.50%		
36) 1,2-Dichloropropane-d6	6.077	67	156889	43.07	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	86.14%		
41) Toluene-d8	7.321	98	445907	41.54	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	83.08%		
43) trans-1,3-Dichloroprop...	7.627	79	69292	38.88	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	77.76%		
47) 2-Hexanone-d5	8.093	63	100400	66.31	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	66.31%		
57) 1,1,2,2-Tetrachloroeth...	10.222	84	194404	38.99	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	77.98%		
64) 1,2-Dichlorobenzene-d4	11.633	152	150697	46.84	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	93.68%		
Target Compounds						Qvalue
20) cis-1,2-Dichloroethene	3.920	96	40172	14.72	ug/L	97
34) Trichloroethene	5.920	95	22089	6.21	ug/L	96
46) Tetrachloroethene	7.981	164	179425	67.53	ug/L	98
48) 2-Hexanone	8.151	43	10856	2.84	ug/L #	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012221\
Data File : VV020137.D
Acq On : 22 Jan 2021 15:51
Operator : SY/MD
Sample : M1181-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
F1E39

Quant Time: Jan 23 00:15:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012121WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jan 23 00:14:09 2021
Response via : Initial Calibration





#20

cis-1,2-Dichloroethene

Concen: 14.72 ug/L

RT: 3.920 min Scan# 880





Delta R.T. 0.006 min

Lab File: VV020137.D

Acq: 22 Jan 2021 15:51

Tgt Ion: 96 Resp: 40172

Ion	Ratio	Lower	Upper
96	100		
61	134.4	96.9	179.9
68	0.0	0.0	0.0



#34

Trichloroethene

Concen: 6.21 ug/L

RT: 5.920 min Scan# 1502




Delta R.T. 0.006 min


Lab File: VV020137.D

Acq: 22 Jan 2021 15:51

Tgt Ion: 95 Resp: 22089

Ion	Ratio	Lower	Upper
95	100		
97	63.8	44.5	82.7
132	96.0	63.4	117.8
130	99.5	67.1	124.5









#46
Tetrachloroethene
Concen: 67.53 ug/L
RT: 7.981 min Scan# 2143
Delta R.T. 0.003 min
Lab File: VV020137.D
Acq: 22 Jan 2021 15:51

Instrument :
MSVOA_V
ClientSampleId :
F1E39

Tgt Ion:164 Resp: 179425

Ion	Ratio	Lower	Upper
164	100		
129	97.5	70.6	131.0
131	92.9	66.8	124.0
166	124.9	88.2	163.8



#48
2-Hexanone
Concen: 2.84 ug/L
RT: 8.151 min Scan# 2196
Delta R.T. 0.009 min
Lab File: VV020137.D
Acq: 22 Jan 2021 15:51

Tgt Ion: 43 Resp: 10856

Ion	Ratio	Lower	Upper
43	100		
58	43.9	41.8	62.6
57	21.7	15.0	22.4
100	6.6	8.9	13.3

