

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012221\
 Data File : VW020144.D
 Acq On : 22 Jan 2021 18:36
 Operator : SY/MD
 Sample : M1201-03
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4Z8

Quant Time: Jan 23 00:17:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jan 23 00:14:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	370635	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.859	117	342523	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.254	152	215858	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.306	65	149653	54.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	108.88%
7) Chloroethane-d5	1.566	69	131237	57.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	114.06%
11) 1,1-Dichloroethene-d2	2.110	63	195989	35.69	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.38%
21) 2-Butanone-d5	3.878	46	191320	101.48	ug/L	-0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.48%
24) Chloroform-d	4.354	84	261303	50.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.50%
26) 1,2-Dichloroethane-d4	5.036	65	185707	52.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.06%
32) Benzene-d6	5.055	84	501006	53.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.22%
36) 1,2-Dichloropropane-d6	6.074	67	179749	60.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	120.72%#
41) Toluene-d8	7.322	98	511367	58.27	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	116.54%
43) trans-1,3-Dichloroprop...	7.624	79	77540	53.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	106.42%
47) 2-Hexanone-d5	8.090	63	122625	99.06	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.06%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	227069	55.70	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	111.40%
64) 1,2-Dichlorobenzene-d4	11.630	152	221286	54.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	109.12%
Target Compounds						
13) Acetone	2.171	43	8838	5.45	ug/L	93
29) Cyclohexane	4.682	56	29109	7.30	ug/L	96
48) 2-Hexanone	8.148	43	10019	3.21	ug/L #	44


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012221\
Data File : VV020144.D
Acq On : 22 Jan 2021 18:36
Operator : SY/MD
Sample : M1201-03
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG4Z8

Quant Time: Jan 23 00:17:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012121WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jan 23 00:14:09 2021
Response via : Initial Calibration









#13
Acetone
Concen: 5.45 ug/L
RT: 2.171 min Scan# 336
Delta R.T. -0.013 min
Lab File: VV020144.D
Acq: 22 Jan 2021 18:36




Instrument :
MSVOA_V
ClientSampled :
BG4Z8


Tgt Ion: 43 Resp: 8838
Ion Ratio Lower Upper
43 100
58 33.6 0.0 60.2



#29
Cyclohexane
Concen: 7.30 ug/L
RT: 4.682 min Scan# 1117
Delta R.T. 0.003 min
Lab File: VV020144.D
Acq: 22 Jan 2021 18:36


Tgt Ion: 56 Resp: 29109
Ion Ratio Lower Upper
56 100
69 30.5 24.2 36.2
84 87.8 66.3 99.5





#48
 2-Hexanone
 Concen: 3.21 ug/L
 RT: 8.148 min Scan# 2195
 Delta R.T. 0.006 min
 Lab File: VV020144.D
 Acq: 22 Jan 2021 18:36

Instrument :
 MSVOA_V
 ClientSampleId :
 BG4Z8



Tgt Ion: 43	Resp: 10019
Ion Ratio	Lower Upper
43	100
58	0.0 41.8 62.6#
57	3.5 15.0 22.4#
100	7.1 8.9 13.3#

