

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012224\
 Data File : VW033799.D
 Acq On : 22 Jan 2024 13:48
 Operator : SY/MD
 Sample : P1190-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMT6

Quant Time: Jan 22 23:44:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 22 23:41:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	223509	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	210665	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	104905	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	75237	43.799	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	87.600%		
7) Chloroethane-d5	1.532	69	61585	45.272	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	90.540%		
11) 1,1-Dichloroethene-d2	2.060	65	34850	47.141	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	94.280%		
21) 2-Butanone-d5	3.783	46	81390	69.078	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	69.080%		
24) Chloroform-d	4.249	84	136012	44.281	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	88.560%		
26) 1,2-Dichloroethane-d4	4.941	65	94236	47.483	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	94.960%		
32) Benzene-d6	4.960	84	274765	45.878	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	91.760%		
36) 1,2-Dichloropropane-d6	5.989	67	86628	43.942	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	87.880%		
41) Toluene-d8	7.243	98	254535	47.169	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	94.340%		
43) trans-1,3-Dichloroprop...	7.555	79	40098	41.155	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	82.320%		
47) 2-Hexanone-d5	8.024	63	62671	76.640	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	76.640%		
56) 1,1,2,2-Tetrachloroeth...	10.153	84	108687	43.185	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	86.380%		
66) 1,2-Dichlorobenzene-d4	11.561	152	96974	48.722	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	97.440%		
Target Compounds						
17) trans-1,2-Dichloroethene	2.703	96	1839	1.227	ug/L	87
20) cis-1,2-Dichloroethene	3.815	96	35617	21.358	ug/L	91
34) Trichloroethene	5.847	95	2315	1.395	ug/L	89
46) Tetrachloroethene	7.912	164	3082	2.532	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012224\
Data File : VV033799.D
Acq On : 22 Jan 2024 13:48
Operator : SY/MD
Sample : P1190-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMT6

Quant Time: Jan 22 23:44:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 22 23:41:48 2024
Response via : Initial Calibration

