

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012224\
 Data File : VV033793.D
 Acq On : 22 Jan 2024 11:25
 Operator : SY/MD
 Sample : VV0122WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK221

Quant Time: Jan 22 23:42:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 22 23:41:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	223626	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	210799	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.184	152	106478	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	74550	43.377	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.760%
7) Chloroethane-d5	1.529	69	63113	46.371	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.740%
11) 1,1-Dichloroethene-d2	2.059	65	34604	46.783	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.560%
21) 2-Butanone-d5	3.783	46	88712	75.253	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	75.250%
24) Chloroform-d	4.246	84	137354	44.694	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.380%
26) 1,2-Dichloroethane-d4	4.940	65	95898	48.295	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.600%
32) Benzene-d6	4.960	84	278201	46.422	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.840%
36) 1,2-Dichloropropane-d6	5.985	67	86519	43.859	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.720%
41) Toluene-d8	7.243	98	256093	47.428	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.860%
43) trans-1,3-Dichloroprop...	7.554	79	42570	43.665	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.320%
47) 2-Hexanone-d5	8.021	63	64939	79.363	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.360%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	107801	42.806	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.620%
66) 1,2-Dichlorobenzene-d4	11.557	152	99525	49.265	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.540%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012224\
Data File : VV033793.D
Acq On : 22 Jan 2024 11:25
Operator : SY/MD
Sample : VV0122WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK221

Quant Time: Jan 22 23:42:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 22 23:41:48 2024
Response via : Initial Calibration

