

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012224\
 Data File : VW033808.D
 Acq On : 22 Jan 2024 17:23
 Operator : SY/MD
 Sample : P1191-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMQ5

Quant Time: Jan 22 23:46:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 22 23:41:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	214975	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	199978	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	100751	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	69337	41.967	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.940%
7) Chloroethane-d5	1.532	69	59833	45.730	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.460%
11) 1,1-Dichloroethene-d2	2.059	65	33161	46.637	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.280%
21) 2-Butanone-d5	3.783	46	81415	71.842	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	71.840%
24) Chloroform-d	4.249	84	131128	44.385	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.780%
26) 1,2-Dichloroethane-d4	4.940	65	90623	47.475	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.960%
32) Benzene-d6	4.960	84	261553	46.006	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.020%
36) 1,2-Dichloropropane-d6	5.989	67	82358	44.009	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.020%
41) Toluene-d8	7.243	98	238248	46.511	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.020%
43) trans-1,3-Dichloroprop...	7.554	79	39506	42.714	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.420%
47) 2-Hexanone-d5	8.021	63	59300	76.393	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.390%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	101457	42.467	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.940%
66) 1,2-Dichlorobenzene-d4	11.561	152	92102	48.182	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012224\
Data File : VV033808.D
Acq On : 22 Jan 2024 17:23
Operator : SY/MD
Sample : P1191-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMQ5

Quant Time: Jan 22 23:46:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 22 23:41:48 2024
Response via : Initial Calibration

