

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012225\
 Data File : VV038504.D
 Acq On : 22 Jan 2025 17:05
 Operator : SY/MD
 Sample : Q1108-18
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK002

Quant Time: Jan 23 01:41:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 23 01:38:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	183383	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	202936	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	83103	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	58798	3.888	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.800%
7) Chloroethane-d5	1.532	69	55781	4.393	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.800%
11) 1,1-Dichloroethene-d2	2.056	65	28186	4.136	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.800%
20) 2-Butanone-d5	3.812	46	134307	42.226	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.460%
24) Chloroform-d	4.240	84	129033	4.552	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
26) 1,2-Dichloroethane-d4	4.937	65	60623	4.667	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.400%
32) Benzene-d6	4.953	84	228266	4.084	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.600%
36) 1,2-Dichloropropane-d6	5.979	67	77158	4.542	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.800%
41) Toluene-d8	7.236	98	176232	3.510	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.200%
43) trans-1,3-Dichloroprop...	7.558	79	23079	3.714	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.200%
46) 2-Hexanone-d5	8.024	63	81981	33.437	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	66.880%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	57861	4.296	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.000%
66) 1,2-Dichlorobenzene-d4	11.555	152	67623	4.808	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012225\
Data File : VV038504.D
Acq On : 22 Jan 2025 17:05
Operator : SY/MD
Sample : Q1108-18
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK002

Quant Time: Jan 23 01:41:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 23 01:38:03 2025
Response via : Initial Calibration

