




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012225\
Data File : VV038502.D
Acq On : 22 Jan 2025 16:19
Operator : SY/MD
Sample : Q1137-05
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHAB8

Manual IntegrationsAPPROVED

Quant Time: Jan 23 01:40:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 23 01:38:03 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/24/2025
Supervised By :Semsettin Yesilyurt 01/24/2025



TIC: VV038502.D\data.ms

(13) Acetone (T)

2.201min (+ 0.042) 8.61 ug/L

response 22484

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	33.60	13.81
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012225\
 Data File : VV038502.D
 Acq On : 22 Jan 2025 16:19
 Operator : SY/MD
 Sample : Q1137-05
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHAB8

Manual IntegrationsAPPROVED

Quant Time: Jan 23 01:40:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725wMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 23 01:38:03 2025
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/24/2025
 Supervised By :Semsettin Yesilyurt 01/24/2025



TIC: VV038502.D\data.ms

(13) Acetone (T)

2.201min (+ 0.042) 17.99 ug/L m

response 46947

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	33.60	6.61
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012225\
 Data File : VV038502.D
 Acq On : 22 Jan 2025 16:19
 Operator : SY/MD
 Sample : Q1137-05
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHAB8

Manual IntegrationsAPPROVED

Quant Time: Jan 23 01:40:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 23 01:38:03 2025
 Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/24/2025
 Supervised By :Semsettin Yesilyurt 01/24/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	193402	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	206974	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	86770	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	56227	3.525	ug/L	0.00
Spiked Amount 5.000	Range 40	- 130	Recovery	=	70.600%	
7) Chloroethane-d5	1.532	69	53188	3.971	ug/L	0.00
Spiked Amount 5.000	Range 65	- 130	Recovery	=	79.400%	
11) 1,1-Dichloroethene-d2	2.056	65	28282	3.935	ug/L	0.00
Spiked Amount 5.000	Range 60	- 125	Recovery	=	78.800%	
20) 2-Butanone-d5	3.860	46	117508	35.030	ug/L	0.04
Spiked Amount 50.000	Range 40	- 130	Recovery	=	70.060%	
24) Chloroform-d	4.243	84	125000	4.182	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125	Recovery	=	83.600%	
26) 1,2-Dichloroethane-d4	4.940	65	56025	4.090	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130	Recovery	=	81.800%	
32) Benzene-d6	4.953	84	226644	3.976	ug/L	0.00
Spiked Amount 5.000	Range 70	- 125	Recovery	=	79.600%	
36) 1,2-Dichloropropane-d6	5.985	67	75137	4.336	ug/L	0.00
Spiked Amount 5.000	Range 60	- 140	Recovery	=	86.800%	
41) Toluene-d8	7.239	98	175269	3.423	ug/L	0.00
Spiked Amount 5.000	Range 70	- 130	Recovery	=	68.400%#	
43) trans-1,3-Dichloroprop...	7.558	79	21523	3.396	ug/L	0.00
Spiked Amount 5.000	Range 55	- 130	Recovery	=	68.000%	
46) 2-Hexanone-d5	8.027	63	81542	32.609	ug/L	0.00
Spiked Amount 50.000	Range 45	- 130	Recovery	=	65.220%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	57210	4.165	ug/L	0.00
Spiked Amount 5.000	Range 65	- 120	Recovery	=	83.200%	
66) 1,2-Dichlorobenzene-d4	11.554	152	72401	4.930	ug/L	0.00
Spiked Amount 5.000	Range 80	- 120	Recovery	=	98.600%	
Target Compounds						
3) Chloromethane	1.214	50	591666	26.692	ug/L	99
13) Acetone	2.201	43	46947m	17.987	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012225\
Data File : VV038502.D
Acq On : 22 Jan 2025 16:19
Operator : SY/MD
Sample : Q1137-05
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHAB8

Manual IntegrationsAPPROVED

Quant Time: Jan 23 01:40:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 23 01:38:03 2025
Response via : Initial Calibration

Reviewed By :Mahesh Dadoda 01/24/2025
Supervised By :Semsettin Yesilyurt 01/24/2025

