

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012323\
 Data File : VW029902.D
 Acq On : 23 Jan 2023 14:38
 Operator : SY/MD
 Sample : VW0123WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK294

Quant Time: Jan 24 01:33:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 24 01:33:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	176229	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	178648	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	95545	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	64836	4.851	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.000%
7) Chloroethane-d5	1.564	69	61106	5.515	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.200%
11) 1,1-Dichloroethene-d2	2.104	63	88083	3.615	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.200%
20) 2-Butanone-d5	3.892	46	129047	54.469	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.940%
24) Chloroform-d	4.339	84	118774	4.844	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	5.024	65	54333	5.232	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.600%
32) Benzene-d6	5.043	84	240630	4.541	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.800%
36) 1,2-Dichloropropane-d6	6.063	67	79774	5.144	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.800%
41) Toluene-d8	7.307	98	198321	4.042	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.800%
43) trans-1,3-Dichloroprop...	7.616	79	23102	4.079	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%
46) 2-Hexanone-d5	8.085	63	108031	46.849	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.700%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	59122	5.202	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.000%
66) 1,2-Dichlorobenzene-d4	11.612	152	79152	4.557	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012323\
Data File : VV029902.D
Acq On : 23 Jan 2023 14:38
Operator : SY/MD
Sample : VV0123WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK294

Quant Time: Jan 24 01:33:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 24 01:33:02 2023
Response via : Initial Calibration

