

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012323\
 Data File : VW029915.D
 Acq On : 23 Jan 2023 20:06
 Operator : SY/MD
 Sample : 01225-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX9G1

Quant Time: Jan 24 01:36:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 24 01:33:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	144024	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	146854	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	71732	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	67096	6.142	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	122.800%
7) Chloroethane-d5	1.564	69	63094	6.967	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	139.400%#
11) 1,1-Dichloroethene-d2	2.101	63	86745	4.356	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.200%
20) 2-Butanone-d5	3.918	46	107037	55.282	ug/L	0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.560%
24) Chloroform-d	4.339	84	117944	5.886	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	117.800%
26) 1,2-Dichloroethane-d4	5.024	65	51410	6.058	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	121.200%
32) Benzene-d6	5.040	84	225734	5.182	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.600%
36) 1,2-Dichloropropane-d6	6.063	67	75041	5.886	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.800%
41) Toluene-d8	7.307	98	179652	4.455	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.000%
43) trans-1,3-Dichloroprop...	7.616	79	19042	4.090	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.800%
46) 2-Hexanone-d5	8.088	63	99396	52.436	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.880%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	54557	5.840	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	116.800%
66) 1,2-Dichlorobenzene-d4	11.612	152	70748	5.425	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012323\
Data File : VV029915.D
Acq On : 23 Jan 2023 20:06
Operator : SY/MD
Sample : 01225-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX9G1

Quant Time: Jan 24 01:36:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 24 01:33:02 2023
Response via : Initial Calibration

