

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012323\
 Data File : VW029916.D
 Acq On : 23 Jan 2023 20:30
 Operator : SY/MD
 Sample : 01225-12
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX9G2

Quant Time: Jan 24 01:37:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 24 01:33:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	147774	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	150837	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	75464	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	64328	5.739	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.800%
7) Chloroethane-d5	1.564	69	61743	6.645	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	133.000%#
11) 1,1-Dichloroethene-d2	2.105	63	83650	4.094	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.800%
20) 2-Butanone-d5	3.899	46	116794	58.790	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.580%
24) Chloroform-d	4.339	84	119017	5.789	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	115.800%
26) 1,2-Dichloroethane-d4	5.027	65	54027	6.205	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	124.000%
32) Benzene-d6	5.043	84	228896	5.116	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.400%
36) 1,2-Dichloropropane-d6	6.063	67	76996	5.880	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	117.600%
41) Toluene-d8	7.310	98	182880	4.415	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%
43) trans-1,3-Dichloroprop...	7.616	79	19402	4.057	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.200%
46) 2-Hexanone-d5	8.088	63	102786	52.792	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.580%
56) 1,1,2,2-Tetrachloroeth...	10.207	84	58333	6.079	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	121.600%#
66) 1,2-Dichlorobenzene-d4	11.612	152	74506	5.431	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012323\
Data File : VV029916.D
Acq On : 23 Jan 2023 20:30
Operator : SY/MD
Sample : 01225-12
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX9G2

Quant Time: Jan 24 01:37:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 24 01:33:02 2023
Response via : Initial Calibration

