

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012323\
 Data File : VW029918.D
 Acq On : 23 Jan 2023 21:18
 Operator : SY/MD
 Sample : 01225-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX9G3

Quant Time: Jan 24 01:37:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 24 01:33:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Difluorobenzene	5.609	114	150944	5.000 ug/L	0.00
28) Chlorobenzene-d5	8.844	117	150275	5.000 ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	78113	5.000 ug/L	0.00
System Monitoring Compounds					
4) Vinyl Chloride-d3	1.304	65	65268	5.701 ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 114.000%	
7) Chloroethane-d5	1.564	69	64841	6.832 ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 136.600%#	
11) 1,1-Dichloroethene-d2	2.104	63	85307	4.087 ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 81.800%	
20) 2-Butanone-d5	3.889	46	125196	61.696 ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 123.400%	
24) Chloroform-d	4.339	84	121752	5.797 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 116.000%	
26) 1,2-Dichloroethane-d4	5.027	65	55647	6.256 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 125.200%	
32) Benzene-d6	5.043	84	231324	5.189 ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 103.800%	
36) 1,2-Dichloropropane-d6	6.063	67	77735	5.959 ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 119.200%	
41) Toluene-d8	7.307	98	178138	4.316 ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 86.400%	
43) trans-1,3-Dichloroprop...	7.615	79	18262	3.833 ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 76.600%	
46) 2-Hexanone-d5	8.085	63	97473	50.251 ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 100.500%	
56) 1,1,2,2-Tetrachloroeth...	10.204	84	58178	6.085 ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 121.800%#	
66) 1,2-Dichlorobenzene-d4	11.612	152	75455	5.313 ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 106.200%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012323\
Data File : VV029918.D
Acq On : 23 Jan 2023 21:18
Operator : SY/MD
Sample : 01225-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX9G3

Quant Time: Jan 24 01:37:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 24 01:33:02 2023
Response via : Initial Calibration

