

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012323\
 Data File : VW029907.D
 Acq On : 23 Jan 2023 16:53
 Operator : SY/MD
 Sample : 01225-01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX9F1

Quant Time: Jan 24 01:35:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 24 01:33:02 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	162981	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	162799	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	85798	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	57423	4.645	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.000%
7) Chloroethane-d5	1.565	69	54499	5.318	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.400%
11) 1,1-Dichloroethene-d2	2.105	63	78954	3.504	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.000%
20) 2-Butanone-d5	3.886	46	120922	55.189	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.380%
24) Chloroform-d	4.339	84	112199	4.948	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.000%
26) 1,2-Dichloroethane-d4	5.024	65	50100	5.217	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.043	84	216273	4.479	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.600%
36) 1,2-Dichloropropane-d6	6.063	67	71111	5.032	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.600%
41) Toluene-d8	7.307	98	177033	3.960	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.200%
43) trans-1,3-Dichloroprop...	7.616	79	20530	3.977	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.085	63	98525	46.886	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.780%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	53869	5.201	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.000%
66) 1,2-Dichlorobenzene-d4	11.612	152	68922	4.418	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012323\
Data File : VV029907.D
Acq On : 23 Jan 2023 16:53
Operator : SY/MD
Sample : 01225-01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX9F1

Quant Time: Jan 24 01:35:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 24 01:33:02 2023
Response via : Initial Calibration

