

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012324\
 Data File : VW033821.D
 Acq On : 23 Jan 2024 11:37
 Operator : SY/MD
 Sample : P1190-13DL 200X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMS4DL

Quant Time: Jan 24 02:02:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 24 02:00:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	230724	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	211416	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.184	152	108409	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	69812	39.370	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	78.740%	
7) Chloroethane-d5	1.532	69	60289	42.933	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	85.860%	
11) 1,1-Dichloroethene-d2	2.063	65	33336	43.683	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	87.360%	
21) 2-Butanone-d5	3.786	46	93316	76.723	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	76.720%	
24) Chloroform-d	4.249	84	135141	42.621	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	85.240%	
26) 1,2-Dichloroethane-d4	4.940	65	93998	45.882	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.760%	
32) Benzene-d6	4.960	84	273598	45.521	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.040%	
36) 1,2-Dichloropropane-d6	5.989	67	84222	42.570	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	85.140%	
41) Toluene-d8	7.243	98	246925	45.597	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	91.200%	
43) trans-1,3-Dichloroprop...	7.554	79	42417	43.381	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	86.760%	
47) 2-Hexanone-d5	8.024	63	64742	78.891	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	78.890%	
56) 1,1,2,2-Tetrachloroeth...	10.152	84	107529	42.574	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	85.140%	
66) 1,2-Dichlorobenzene-d4	11.561	152	96422	46.879	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	93.760%	
Target Compounds						
20) cis-1,2-Dichloroethene	3.821	96	11973	6.955	ug/L	95
34) Trichloroethene	5.837	95	20312	12.194	ug/L	96
46) Tetrachloroethene	7.905	164	58376	47.780	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012324\
Data File : VV033821.D
Acq On : 23 Jan 2024 11:37
Operator : SY/MD
Sample : P1190-13DL 200X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMS4DL

Quant Time: Jan 24 02:02:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 24 02:00:55 2024
Response via : Initial Calibration

