

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012324\
 Data File : VW033837.D
 Acq On : 23 Jan 2024 18:22
 Operator : SY/MD
 Sample : P1191-13 50X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMQ3

Quant Time: Jan 24 02:06:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jan 24 02:00:55 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	220754	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	203245	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	100391	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	55577	32.758	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	65.520%
7) Chloroethane-d5	1.532	69	50824	37.827	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.660%
11) 1,1-Dichloroethene-d2	2.060	65	28030	38.389	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.780%
21) 2-Butanone-d5	3.783	46	76069	65.368	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	65.370%
24) Chloroform-d	4.249	84	117408	38.701	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.400%
26) 1,2-Dichloroethane-d4	4.941	65	82000	41.833	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.660%
32) Benzene-d6	4.960	84	230203	39.840	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.680%
36) 1,2-Dichloropropane-d6	5.986	67	74123	38.972	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	77.940%
41) Toluene-d8	7.243	98	207768	39.908	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.820%#
43) trans-1,3-Dichloroprop...	7.555	79	35027	37.263	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.520%
47) 2-Hexanone-d5	8.024	63	53304	67.565	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	67.560%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	94593	38.957	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.920%
66) 1,2-Dichlorobenzene-d4	11.561	152	81229	42.647	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.300%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012324\
Data File : VV033837.D
Acq On : 23 Jan 2024 18:22
Operator : SY/MD
Sample : P1191-13 50X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMQ3

Quant Time: Jan 24 02:06:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jan 24 02:00:55 2024
Response via : Initial Calibration

