

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012325\
 Data File : VV038511.D
 Acq On : 23 Jan 2025 17:12
 Operator : SY/MD
 Sample : VV0123WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK276

Quant Time: Jan 24 00:31:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 24 00:30:49 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	194250	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	204879	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	83184	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	48975	3.057	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.200%
7) Chloroethane-d5	1.532	69	47319	3.518	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.400%
11) 1,1-Dichloroethene-d2	2.053	65	23496	3.255	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.000%
20) 2-Butanone-d5	3.844	46	127487	37.839	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.680%
24) Chloroform-d	4.240	84	114033	3.798	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.000%
26) 1,2-Dichloroethane-d4	4.941	65	54712	3.977	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.600%
32) Benzene-d6	4.953	84	202175	3.583	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	71.600%
36) 1,2-Dichloropropane-d6	5.982	67	67427	3.931	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.600%
41) Toluene-d8	7.240	98	157148	3.100	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	62.000%#
43) trans-1,3-Dichloroprop...	7.561	79	20071	3.199	ug/L	0.01
Spiked Amount	5.000	Range	55 - 130	Recovery	=	64.000%
46) 2-Hexanone-d5	8.027	63	76591	30.943	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	61.880%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	54606	4.016	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.400%
66) 1,2-Dichlorobenzene-d4	11.551	152	61941	4.399	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012325\
Data File : VV038511.D
Acq On : 23 Jan 2025 17:12
Operator : SY/MD
Sample : VV0123WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK276

Quant Time: Jan 24 00:31:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 24 00:30:49 2025
Response via : Initial Calibration

