

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012423\
 Data File : VW029927.D
 Acq On : 24 Jan 2023 16:52
 Operator : SY/MD
 Sample : 01225-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EX9G0

Quant Time: Jan 25 00:45:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 25 00:44:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	178704	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	171435	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.239	152	84445	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	68353	5.043	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	100.800%	
7) Chloroethane-d5	1.564	69	65681	5.846	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	117.000%	
11) 1,1-Dichloroethene-d2	2.101	63	92048	3.725	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	74.600%	
20) 2-Butanone-d5	3.915	46	104672	43.569	ug/L	0.03
Spiked Amount	50.000	Range 40 - 130	Recovery	=	87.140%	
24) Chloroform-d	4.339	84	122228	4.916	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.400%	
26) 1,2-Dichloroethane-d4	5.024	65	55087	5.231	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.600%	
32) Benzene-d6	5.040	84	243347	4.785	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.800%	
36) 1,2-Dichloropropane-d6	6.063	67	79509	5.342	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	106.800%	
41) Toluene-d8	7.307	98	196106	4.165	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.400%	
43) trans-1,3-Dichloroprop...	7.616	79	20889	3.843	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	76.800%	
46) 2-Hexanone-d5	8.088	63	102668	46.396	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	92.800%	
56) 1,1,2,2-Tetrachloroeth...	10.204	84	55514	5.090	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.800%	
66) 1,2-Dichlorobenzene-d4	11.612	152	74874	4.877	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	97.600%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.134	85	2834	0.170	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012423\
Data File : VV029927.D
Acq On : 24 Jan 2023 16:52
Operator : SY/MD
Sample : 01225-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EX9G0

Quant Time: Jan 25 00:45:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 25 00:44:28 2023
Response via : Initial Calibration

