

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012425\
 Data File : VV038525.D
 Acq On : 24 Jan 2025 10:03
 Operator : SY/MD
 Sample : VV0124WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK278

Quant Time: Jan 24 20:52:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 24 20:51:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	176106	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	194889	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	78135	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	71431	4.918	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.400%
7) Chloroethane-d5	1.532	69	62962	5.163	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.200%
11) 1,1-Dichloroethene-d2	2.056	65	33010	5.044	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.800%
20) 2-Butanone-d5	3.821	46	122779	40.196	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.400%
24) Chloroform-d	4.243	84	140023	5.144	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
26) 1,2-Dichloroethane-d4	4.937	65	65644	5.263	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.200%
32) Benzene-d6	4.953	84	254760	4.746	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
36) 1,2-Dichloropropane-d6	5.982	67	84052	5.152	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.000%
41) Toluene-d8	7.236	98	204732	4.246	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.000%
43) trans-1,3-Dichloroprop...	7.554	79	26232	4.395	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.000%
46) 2-Hexanone-d5	8.024	63	66467	28.229	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	56.460%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	59687	4.615	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	11.551	152	72840	5.508	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012425\
Data File : VV038525.D
Acq On : 24 Jan 2025 10:03
Operator : SY/MD
Sample : VV0124WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK278

Quant Time: Jan 24 20:52:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 24 20:51:39 2025
Response via : Initial Calibration

