

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012425\
 Data File : VV038529.D
 Acq On : 24 Jan 2025 11:52
 Operator : SY/MD
 Sample : VIBLK213
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK213

Quant Time: Jan 24 20:53:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 24 20:51:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	206847	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	214659	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	90145	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	74533	4.369	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.400%
7) Chloroethane-d5	1.532	69	66303	4.629	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.600%
11) 1,1-Dichloroethene-d2	2.057	65	35814	4.659	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.200%
20) 2-Butanone-d5	3.822	46	139531	38.892	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.780%
24) Chloroform-d	4.240	84	147577	4.616	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
26) 1,2-Dichloroethane-d4	4.937	65	67268	4.592	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.800%
32) Benzene-d6	4.950	84	279909	4.734	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
36) 1,2-Dichloropropane-d6	5.979	67	87473	4.867	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.400%
41) Toluene-d8	7.236	98	219573	4.135	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
43) trans-1,3-Dichloroprop...	7.558	79	28620	4.354	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.000%
46) 2-Hexanone-d5	8.024	63	77367	29.832	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	59.660%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	64119	4.501	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.000%
66) 1,2-Dichlorobenzene-d4	11.551	152	79210	5.192	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012425\
Data File : VV038529.D
Acq On : 24 Jan 2025 11:52
Operator : SY/MD
Sample : VIBLK213
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK213

Quant Time: Jan 24 20:53:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 24 20:51:39 2025
Response via : Initial Calibration

