

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012425\
 Data File : VV038535.D
 Acq On : 24 Jan 2025 16:02
 Operator : SY/MD
 Sample : VIBLK215
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK215

Quant Time: Jan 24 20:54:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 24 20:51:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	257567	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	267679	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	109924	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	70285	3.309	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	66.200%
7) Chloroethane-d5	1.529	69	62835	3.523	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	70.400%
11) 1,1-Dichloroethene-d2	2.053	65	33166	3.465	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.400%
20) 2-Butanone-d5	3.818	46	138144	30.923	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	61.840%
24) Chloroform-d	4.240	84	143289	3.599	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	72.000%
26) 1,2-Dichloroethane-d4	4.934	65	64356	3.528	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.600%
32) Benzene-d6	4.950	84	260807	3.538	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	70.800%
36) 1,2-Dichloropropane-d6	5.979	67	84043	3.750	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.000%
41) Toluene-d8	7.236	98	206823	3.123	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	62.400%#
43) trans-1,3-Dichloroprop...	7.558	79	24944	3.043	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	60.800%
46) 2-Hexanone-d5	8.024	63	71961	22.252	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	44.500%#
56) 1,1,2,2-Tetrachloroeth...	10.146	84	60188	3.388	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	67.800%
66) 1,2-Dichlorobenzene-d4	11.554	152	76464	4.110	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012425\
Data File : VV038535.D
Acq On : 24 Jan 2025 16:02
Operator : SY/MD
Sample : VIBLK215
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK215

Quant Time: Jan 24 20:54:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 24 20:51:39 2025
Response via : Initial Calibration

