

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012523\
 Data File : VW029937.D
 Acq On : 25 Jan 2023 12:39
 Operator : SY/MD
 Sample : 01225-13
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jan 25 22:13:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 25 22:12:17 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	172990	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.844	117	173139	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	96605	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	50801	3.872	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.400%
7) Chloroethane-d5	1.564	69	52581	4.834	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.600%
11) 1,1-Dichloroethene-d2	2.101	63	71696	2.997	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.000%
20) 2-Butanone-d5	3.889	46	140570	60.444	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.880%
24) Chloroform-d	4.336	84	111427	4.629	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	5.024	65	49504	4.856	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.200%
32) Benzene-d6	5.040	84	203536	3.963	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.200%
36) 1,2-Dichloropropane-d6	6.059	67	71685	4.769	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	7.307	98	163469	3.438	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.800%#
43) trans-1,3-Dichloroprop...	7.616	79	18256	3.326	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	66.600%
46) 2-Hexanone-d5	8.085	63	103311	46.227	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.460%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	56847	5.161	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.200%
66) 1,2-Dichlorobenzene-d4	11.612	152	68312	3.889	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	77.800%#
Target Compounds						Qvalue
16) Methylene chloride	2.500	84	5239	0.286	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012523\
Data File : VV029937.D
Acq On : 25 Jan 2023 12:39
Operator : SY/MD
Sample : 01225-13
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jan 25 22:13:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 25 22:12:17 2023
Response via : Initial Calibration

