

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012523\
 Data File : VW029949.D
 Acq On : 25 Jan 2023 17:53
 Operator : SY/MD
 Sample : 01292-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COL63

Quant Time: Jan 25 22:15:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Jan 25 22:12:17 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards							
1) 1,4-Difluorobenzene		5.609	114	241478	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.844	117	237421	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.236	152	115544	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.304	65	100740	5.500	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.000%	
7) Chloroethane-d5		1.564	69	92189	6.072	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.400%	
11) 1,1-Dichloroethene-d2		2.105	63	137681	4.123	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.400%	
20) 2-Butanone-d5		3.886	46	163568	50.385	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.780%	
24) Chloroform-d		4.339	84	163564	4.868	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.400%	
26) 1,2-Dichloroethane-d4		5.024	65	71970	5.058	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%	
32) Benzene-d6		5.043	84	340638	4.837	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%	
36) 1,2-Dichloropropane-d6		6.059	67	107866	5.233	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.600%	
41) Toluene-d8		7.307	98	287459	4.409	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.200%	
43) trans-1,3-Dichloroprop...		7.616	79	30729	4.082	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.600%	
46) 2-Hexanone-d5		8.085	63	141803	46.271	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.540%	
56) 1,1,2,2-Tetrachloroeth...		10.204	84	72151	4.777	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.600%	
66) 1,2-Dichlorobenzene-d4		11.612	152	101848	4.848	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.000%	
Target Compounds							Qvalue
13) Acetone		2.175	43	4415	1.905	ug/L	93
17) Methyl tert-butyl Ether		2.767	73	16616	0.466	ug/L	95
34) Trichloroethene		5.908	95	53202	2.674	ug/L	97
42) Toluene		7.381	91	40640	0.480	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012523\
Data File : VV029949.D
Acq On : 25 Jan 2023 17:53
Operator : SY/MD
Sample : 01292-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COL63

Quant Time: Jan 25 22:15:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Jan 25 22:12:17 2023
Response via : Initial Calibration

