

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012524\
 Data File : VW033864.D
 Acq On : 25 Jan 2024 12:39
 Operator : SY/MD
 Sample : VW0125WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK399

Quant Time: Jan 25 21:19:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 25 21:12:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	100909	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	92393	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	45294	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	39618	5.069	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.400%
7) Chloroethane-d5	1.532	69	35110	5.129	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.600%
11) 1,1-Dichloroethene-d2	2.057	65	19566	5.191	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	103.800%
20) 2-Butanone-d5	3.818	46	48212	36.732	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	73.460%
24) Chloroform-d	4.249	84	73087	5.147	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.000%
26) 1,2-Dichloroethane-d4	4.944	65	35088	5.287	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.800%
32) Benzene-d6	4.960	84	147443	5.319	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.400%
36) 1,2-Dichloropropane-d6	5.989	67	42248	5.452	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.000%
41) Toluene-d8	7.243	98	135935	5.336	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
43) trans-1,3-Dichloroprop...	7.558	79	15260	5.216	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	104.400%
46) 2-Hexanone-d5	8.027	63	52702	50.458	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.920%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	25047	5.332	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	44380	5.388	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012524\
Data File : VV033864.D
Acq On : 25 Jan 2024 12:39
Operator : SY/MD
Sample : VV0125WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK399

Quant Time: Jan 25 21:19:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 25 21:12:54 2024
Response via : Initial Calibration

