

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012524\
 Data File : VV033865.D
 Acq On : 25 Jan 2024 13:03
 Operator : SY/MD
 Sample : VV0125WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK200

Quant Time: Jan 25 21:19:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 25 21:12:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	99659	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	92253	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	44703	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	37866	4.906	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.200%
7) Chloroethane-d5	1.532	69	34549	5.110	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.200%
11) 1,1-Dichloroethene-d2	2.060	65	18224	4.896	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.000%
20) 2-Butanone-d5	3.822	46	48213	37.194	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	74.380%
24) Chloroform-d	4.253	84	68861	4.910	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	4.947	65	34070	5.198	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.000%
32) Benzene-d6	4.960	84	140009	5.058	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.200%
36) 1,2-Dichloropropane-d6	5.992	67	40145	5.189	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.800%
41) Toluene-d8	7.246	98	128458	5.050	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.000%
43) trans-1,3-Dichloroprop...	7.561	79	14052	4.810	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	96.200%
46) 2-Hexanone-d5	8.031	63	53249	51.059	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.120%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	24214	5.162	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	43129	5.305	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012524\
Data File : VV033865.D
Acq On : 25 Jan 2024 13:03
Operator : SY/MD
Sample : VV0125WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK200

Quant Time: Jan 25 21:19:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Jan 25 21:12:54 2024
Response via : Initial Calibration

