

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW012622\  
 Data File : VW024449.D  
 Acq On : 26 Jan 2022 19:06  
 Operator : SY/MD  
 Sample : N1275-16  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 C0F53

Quant Time: Jan 27 03:52:41 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR012622WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Jan 27 01:56:16 2022  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	154422	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	168930	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	83507	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	67613	4.998	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	100.000%	
7) Chloroethane-d5	1.571	69	63153	5.351	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	107.000%	
11) 1,1-Dichloroethene-d2	2.111	63	101719	3.934	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	78.600%	
20) 2-Butanone-d5	3.902	46	132029	61.233	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	122.460%	
24) Chloroform-d	4.352	84	139326	5.427	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	108.600%	
26) 1,2-Dichloroethane-d4	5.034	65	64788	5.572	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	111.400%	
32) Benzene-d6	5.053	84	256799	5.090	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.800%	
36) 1,2-Dichloropropane-d6	6.069	67	82233	5.176	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	103.600%	
41) Toluene-d8	7.317	98	215492	4.655	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.200%	
43) trans-1,3-Dichloroprop...	7.625	79	25114	4.603	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	92.000%	
46) 2-Hexanone-d5	8.088	63	88582	47.487	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.980%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	47212	4.993	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	99.800%	
66) 1,2-Dichlorobenzene-d4	11.625	152	80547	5.625	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	112.600%	
Target Compounds						
25) Chloroform	4.381	83	13386	0.568	ug/L	91
-----						

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV012622\  
 Data File : VV024449.D  
 Acq On : 26 Jan 2022 19:06  
 Operator : SY/MD  
 Sample : N1275-16  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 C0F53

Quant Time: Jan 27 03:52:41 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR012622WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Jan 27 01:56:16 2022  
 Response via : Initial Calibration

