

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012623\
 Data File : VV029961.D
 Acq On : 26 Jan 2023 14:10
 Operator : SY/MD
 Sample : VIBLK272
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK272

Quant Time: Jan 27 01:30:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 27 01:28:41 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.606	114	251610	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.841	117	239279	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.236	152	123607	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.304	65	99803	5.230	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.600%	
7) Chloroethane-d5		1.564	69	90006	5.689	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.800%	
11) 1,1-Dichloroethene-d2		2.105	63	132249	3.801	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.000%	
20) 2-Butanone-d5		3.883	46	171450	50.686	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.380%	
24) Chloroform-d		4.336	84	161232	4.606	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%	
26) 1,2-Dichloroethane-d4		5.021	65	69900	4.715	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.200%	
32) Benzene-d6		5.040	84	343838	4.844	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%	
36) 1,2-Dichloropropane-d6		6.059	67	107475	5.174	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.400%	
41) Toluene-d8		7.304	98	294849	4.487	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.800%	
43) trans-1,3-Dichloroprop...		7.609	79	34708	4.575	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.400%	
46) 2-Hexanone-d5		8.082	63	140905	45.621	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.240%	
56) 1,1,2,2-Tetrachloroeth...		10.204	84	72747	4.779	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.600%	
66) 1,2-Dichlorobenzene-d4		11.612	152	105353	4.688	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.800%	
Target Compounds							
13) Acetone		2.182	43	15341	6.354	ug/L	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012623\
Data File : VV029961.D
Acq On : 26 Jan 2023 14:10
Operator : SY/MD
Sample : VIBLK272
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK272

Quant Time: Jan 27 01:30:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 27 01:28:41 2023
Response via : Initial Calibration

