

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012624\
 Data File : VW033873.D
 Acq On : 26 Jan 2024 09:31
 Operator : SY/MD
 Sample : VW0126MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK224

Quant Time: Jan 29 02:00:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 29 01:58:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	218366	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	204571	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	105118	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	76597	45.641	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.280%
7) Chloroethane-d5	1.516	69	60324	45.389	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.780%
11) 1,1-Dichloroethene-d2	2.053	65	35145	48.659	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.320%
21) 2-Butanone-d5	3.783	46	88743	77.093	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.090%
24) Chloroform-d	4.246	84	129227	43.063	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.120%
26) 1,2-Dichloroethane-d4	4.937	65	89879	46.354	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.700%
32) Benzene-d6	4.957	84	254217	43.711	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.420%
36) 1,2-Dichloropropane-d6	5.985	67	78412	40.960	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	81.920%
41) Toluene-d8	7.243	98	238937	45.598	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.200%
43) trans-1,3-Dichloroprop...	7.551	79	41843	44.226	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.460%
47) 2-Hexanone-d5	8.021	63	70564	88.863	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	88.860%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	105514	43.174	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.340%
66) 1,2-Dichlorobenzene-d4	11.561	152	93663	46.963	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012624\
Data File : VV033873.D
Acq On : 26 Jan 2024 09:31
Operator : SY/MD
Sample : VV0126MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK224

Quant Time: Jan 29 02:00:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 29 01:58:36 2024
Response via : Initial Calibration

