

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012624\
 Data File : VW033887.D
 Acq On : 26 Jan 2024 15:09
 Operator : SY/MD
 Sample : P1252-05 50X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 DCMS6

Quant Time: Jan 29 02:03:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 29 01:58:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	210408	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	196467	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	96825	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	67554	41.775	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.560%
7) Chloroethane-d5	1.529	69	53918	42.103	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.200%
11) 1,1-Dichloroethene-d2	2.060	65	31459	45.203	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.400%
21) 2-Butanone-d5	3.783	46	87028	78.462	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	78.460%
24) Chloroform-d	4.249	84	117417	40.607	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.220%
26) 1,2-Dichloroethane-d4	4.941	65	82642	44.234	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.460%
32) Benzene-d6	4.960	84	230798	41.321	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.640%
36) 1,2-Dichloropropane-d6	5.989	67	71855	39.083	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	78.160%
41) Toluene-d8	7.243	98	207835	41.299	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.600%
43) trans-1,3-Dichloroprop...	7.555	79	35925	39.537	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.080%
47) 2-Hexanone-d5	8.024	63	65604	86.024	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.020%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	98068	41.782	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.560%
66) 1,2-Dichlorobenzene-d4	11.561	152	81526	44.379	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.760%
Target Compounds						
20) cis-1,2-Dichloroethene	3.825	96	2382	1.517	ug/L	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012624\
Data File : VV033887.D
Acq On : 26 Jan 2024 15:09
Operator : SY/MD
Sample : P1252-05 50X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
DCMS6

Quant Time: Jan 29 02:03:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 29 01:58:36 2024
Response via : Initial Calibration

