

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012624\
 Data File : VW033874.D
 Acq On : 26 Jan 2024 09:55
 Operator : SY/MD
 Sample : VW0126WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK225

Quant Time: Jan 29 02:00:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 29 01:58:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	210770	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	196664	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	98319	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	77022	47.548	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.100%
7) Chloroethane-d5	1.529	69	60964	47.524	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.040%
11) 1,1-Dichloroethene-d2	2.060	65	34419	49.372	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.740%
21) 2-Butanone-d5	3.783	46	96308	86.680	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.680%
24) Chloroform-d	4.249	84	129679	44.771	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.540%
26) 1,2-Dichloroethane-d4	4.941	65	90042	48.112	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.220%
32) Benzene-d6	4.960	84	256940	45.956	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.920%
36) 1,2-Dichloropropane-d6	5.989	67	78820	42.828	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.660%
41) Toluene-d8	7.243	98	231596	45.974	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.940%
43) trans-1,3-Dichloroprop...	7.555	79	39844	43.806	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.620%
47) 2-Hexanone-d5	8.024	63	72254	94.649	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.650%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	106917	45.507	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.020%
66) 1,2-Dichlorobenzene-d4	11.561	152	90051	48.275	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.540%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012624\
Data File : VV033874.D
Acq On : 26 Jan 2024 09:55
Operator : SY/MD
Sample : VV0126WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK225

Quant Time: Jan 29 02:00:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 29 01:58:36 2024
Response via : Initial Calibration

