

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012624\
 Data File : VW033875.D
 Acq On : 26 Jan 2024 10:21
 Operator : SY/MD
 Sample : PB158608TB 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VLEB608

Quant Time: Jan 29 02:00:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Mon Jan 29 01:58:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	219454	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	204716	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	103023	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	80765	47.886	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.780%
7) Chloroethane-d5	1.529	69	64032	47.940	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.880%
11) 1,1-Dichloroethene-d2	2.060	65	36452	50.219	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.440%
21) 2-Butanone-d5	3.783	46	88172	76.217	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	76.220%
24) Chloroform-d	4.249	84	134907	44.733	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.460%
26) 1,2-Dichloroethane-d4	4.940	65	92819	47.633	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.260%
32) Benzene-d6	4.960	84	265972	45.700	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	5.989	67	83470	43.571	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.140%
41) Toluene-d8	7.243	98	246966	47.097	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.200%
43) trans-1,3-Dichloroprop...	7.555	79	41243	43.561	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.120%
47) 2-Hexanone-d5	8.021	63	67916	85.467	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.470%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	105778	43.251	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.500%
66) 1,2-Dichlorobenzene-d4	11.561	152	95472	48.844	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012624\
Data File : VV033875.D
Acq On : 26 Jan 2024 10:21
Operator : SY/MD
Sample : PB158608TB 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VLEB608

Quant Time: Jan 29 02:00:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM011024WMA.M
Quant Title : VOC Analysis
QLast Update : Mon Jan 29 01:58:36 2024
Response via : Initial Calibration

