

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV012720\
 Data File : VV014408.D
 Acq On : 27 Jan 2020 16:14
 Operator : SY/MD
 Sample : L1253-06DL 5X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AC3DL

Quant Time: Jan 28 05:02:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 03:12:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	874054	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	805294	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	374755	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	190918	3.12	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.40%
7) Chloroethane-d5	1.59	69	184169	3.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.00%
11) 1,1-Dichloroethene-d2	2.13	63	230558	2.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.20%#
20) 2-Butanone-d5	3.93	46	874177	59.50	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	119.00%
24) Chloroform-d	4.40	84	561004	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.00%
26) 1,2-Dichloroethane-d4	5.08	65	277202	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	5.10	84	1048051	4.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.20%
36) 1,2-Dichloropropane-d6	6.12	67	360821	4.84	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.80%
41) Toluene-d8	7.36	98	825767	3.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.20%
43) trans-1,3-Dichloropropene-	7.66	79	127933	3.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.60%
46) 2-Hexanone-d5	8.13	63	804387	57.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.28%
57) 1,1,2,2-Tetrachloroethane-	10.25	84	285572	5.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.60%
64) 1,2-Dichlorobenzene-d4	11.67	152	328472	4.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.40%

Target Compounds

					Ovalue
13) Acetone	2.19	43	34537	3.414 ug/L	90
18) trans-1,2-Dichloroethene	2.79	96	27050	0.461 ug/L	99
19) 1,1-Dichloroethane	3.23	63	42647	0.371 ug/L	97
22) cis-1,2-Dichloroethene	3.96	96	5276683	82.005 ug/L #	98
25) Chloroform	4.43	83	54225	0.492 ug/L	100
29) 1,1,1-Trichloroethane	4.66	97	45938	0.479 ug/L	97
34) Trichloroethene	5.96	95	4420196	69.355 ug/L	98
45) 1,1,2-Trichloroethane	7.88	97	9579	0.240 ug/L	91
47) Tetrachloroethene	8.02	164	8724	0.200 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV012720\
Data File : VV014408.D
Acq On : 27 Jan 2020 16:14
Operator : SY/MD
Sample : L1253-06DL 5X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AC3DL

Quant Time: Jan 28 05:02:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 03:12:38 2020
Response via : Initial Calibration

