

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV012720\
 Data File : VV014410.D
 Acq On : 27 Jan 2020 17:50
 Operator : SY/MD
 Sample : L1253-06DL 5X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AC3DL

Quant Time: Jan 28 11:10:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 03:12:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	888627	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	829730	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	388345	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	215594	3.47	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.40%
7) Chloroethane-d5	1.59	69	186173	3.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.40%
11) 1,1-Dichloroethene-d2	2.13	63	227296	2.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	49.60%#
20) 2-Butanone-d5	3.93	46	861633	57.69	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	115.38%
24) Chloroform-d	4.40	84	534641	4.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.00%
26) 1,2-Dichloroethane-d4	5.08	65	266610	4.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.60%
32) Benzene-d6	5.10	84	1011769	4.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.80%
36) 1,2-Dichloropropane-d6	6.12	67	355362	4.63	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.60%
41) Toluene-d8	7.36	98	798304	3.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.60%
43) trans-1,3-Dichloropropene-	7.66	79	131737	3.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.60%
46) 2-Hexanone-d5	8.14	63	757446	52.68	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.36%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	279571	5.21	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.20%
64) 1,2-Dichlorobenzene-d4	11.67	152	325674	4.47	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.40%

Target Compounds

					Ovalue
13) Acetone	2.19	43	25237	2.454 ug/L	95
16) Methylene chloride	2.54	84	14698	0.227 ug/L	95
22) cis-1,2-Dichloroethene	3.96	96	954635	14.593 ug/L #	96
34) Trichloroethene	5.96	95	481305	7.330 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV012720\
Data File : VV014410.D
Acq On : 27 Jan 2020 17:50
Operator : SY/MD
Sample : L1253-06DL 5X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AC3DL

Quant Time: Jan 28 11:10:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 03:12:38 2020
Response via : Initial Calibration

