

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012722\
 Data File : VW024472.D
 Acq On : 27 Jan 2022 23:35
 Operator : SY/MD
 Sample : N1317-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0AA4

Quant Time: Jan 28 03:24:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Jan 28 03:20:33 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	137740	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	152016	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	72679	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	52424	4.344	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	86.800%	
7) Chloroethane-d5	1.568	69	49756	4.727	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	94.600%	
11) 1,1-Dichloroethene-d2	2.108	63	80323	3.482	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.600%	
20) 2-Butanone-d5	3.918	46	83374	43.351	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	86.700%	
24) Chloroform-d	4.349	84	115799	5.057	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.200%	
26) 1,2-Dichloroethane-d4	5.037	65	56703	5.468	ug/L	0.01
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.400%	
32) Benzene-d6	5.050	84	213192	4.696	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	94.000%	
36) 1,2-Dichloropropane-d6	6.069	67	69319	4.849	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.000%	
41) Toluene-d8	7.317	98	175557	4.214	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.200%	
43) trans-1,3-Dichloroprop...	7.625	79	21629	4.406	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	88.200%	
46) 2-Hexanone-d5	8.092	63	81428	48.509	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.020%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	44410	5.219	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	104.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	67070	5.382	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	107.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012722\
Data File : VV024472.D
Acq On : 27 Jan 2022 23:35
Operator : SY/MD
Sample : N1317-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0AA4

Quant Time: Jan 28 03:24:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Jan 28 03:20:33 2022
Response via : Initial Calibration

