

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012723\
 Data File : VW029977.D
 Acq On : 27 Jan 2023 13:59
 Operator : SY/MD
 Sample : 01315-05
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Jan 28 02:24:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Jan 28 02:00:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.606	114	218972	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.841	117	210427	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.236	152	105084	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	83574	5.032	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.600%
7) Chloroethane-d5	1.564	69	77659	5.641	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	112.800%
11) 1,1-Dichloroethene-d2	2.105	63	112692	3.722	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%
20) 2-Butanone-d5	3.886	46	149847	50.903	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.800%
24) Chloroform-d	4.336	84	141170	4.634	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.600%
26) 1,2-Dichloroethane-d4	5.024	65	61795	4.789	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.800%
32) Benzene-d6	5.040	84	301872	4.836	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
36) 1,2-Dichloropropane-d6	6.059	67	93807	5.135	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.800%
41) Toluene-d8	7.307	98	252482	4.369	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	7.612	79	29244	4.383	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.600%
46) 2-Hexanone-d5	8.082	63	131147	48.284	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.560%
56) 1,1,2,2-Tetrachloroeth...	10.204	84	64143	4.791	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.800%
66) 1,2-Dichlorobenzene-d4	11.612	152	92704	4.852	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012723\
Data File : VV029977.D
Acq On : 27 Jan 2023 13:59
Operator : SY/MD
Sample : 01315-05
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Jan 28 02:24:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011223WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Jan 28 02:00:30 2023
Response via : Initial Calibration

