

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012725\
 Data File : VV038545.D
 Acq On : 27 Jan 2025 10:16
 Operator : SY/MD
 Sample : VSTD00236
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002236

Quant Time: Jan 28 08:01:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:00:57 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6277	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	6215	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2870	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	12254	1.885	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.946	65	9827	2.300	ug/L	-0.02
7) 1,2-Dichloropropane-d6	5.972	67	13220	2.475	ug/L	-0.02
8) Toluene-d8	7.238	98	26617	2.085	ug/L	-0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	10258	2.338	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	16059	2.360	ug/L	98
6) Trichloroethene	5.829	95	10960	2.349	ug/L	98
9) 1,2-Dibromoethane	8.290	107	7513	2.294	ug/L #	96
12) 1,2,3-Trichloropropane	10.197	75	6730	2.509	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.357	75	1067	2.109	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012725\
Data File : VV038545.D
Acq On : 27 Jan 2025 10:16
Operator : SY/MD
Sample : VSTD00236
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002236

Quant Time: Jan 28 08:01:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:00:57 2025
Response via : Initial Calibration

