

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012725\
 Data File : VV038546.D
 Acq On : 27 Jan 2025 10:47
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV237

Quant Time: Jan 28 08:08:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5585	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.801	117	5427	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.192	152	2538	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.281	65	3034	0.539	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 108.000%			
4) 1,2-Dichloroethane-d4	4.970	65	2424	0.539	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 108.000%			
7) 1,2-Dichloropropane-d6	5.996	67	2987	0.560	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 112.000%			
8) Toluene-d8	7.254	98	5994	0.561	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 112.000%			
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2440	0.536	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 108.000%			
Target Compounds					Qvalue	
3) Vinyl chloride	1.281	62	3998	0.550	ug/L	99
6) Trichloroethene	5.853	95	2604	0.581	ug/L	99
9) 1,2-Dibromoethane	8.306	107	1712	0.538	ug/L	100
12) 1,2,3-Trichloropropane	10.198	75	1536	0.529	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.374	75	246	0.538	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012725\
Data File : VV038546.D
Acq On : 27 Jan 2025 10:47
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV237

Quant Time: Jan 28 08:08:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

