

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV012820\
 Data File : VV014421.D
 Acq On : 28 Jan 2020 16:05
 Operator : SY/MD
 Sample : L1271-21
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Jan 29 05:55:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Jan 29 05:25:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	789655	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	728531	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	328303	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	235622	4.27	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.40%
7) Chloroethane-d5	1.59	69	196783	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.20%
11) 1,1-Dichloroethene-d2	2.13	63	281968	3.46	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.20%
20) 2-Butanone-d5	3.93	46	676670	50.98	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.96%
24) Chloroform-d	4.40	84	554735	5.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.80%
26) 1,2-Dichloroethane-d4	5.08	65	262688	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.20%
32) Benzene-d6	5.10	84	1174983	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.00%
36) 1,2-Dichloropropane-d6	6.11	67	378745	5.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.40%
41) Toluene-d8	7.36	98	1077012	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.40%
43) trans-1,3-Dichloropropene-	7.66	79	139303	4.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.40%
46) 2-Hexanone-d5	8.14	63	592854	46.96	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.92%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	255310	5.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	355225	5.76	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.20%

Target Compounds					Ovalue
13) Acetone	2.19	43	20425	2.235 ug/L	98
16) Methylene chloride	2.54	84	24155	0.421 ug/L	97
25) Chloroform	4.42	83	21456	0.216 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV012820\
Data File : VV014421.D
Acq On : 28 Jan 2020 16:05
Operator : SY/MD
Sample : L1271-21
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Jan 29 05:55:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR010220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Jan 29 05:25:35 2020
Response via : Initial Calibration

