

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012821\
 Data File : VV020190.D
 Acq On : 27 Jan 2021 14:28
 Operator : SY/MD
 Sample : VV0128WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK89

Quant Time: Jan 28 05:48:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 28 05:47:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	945379	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.855	117	910958	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	484086	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.309	65	343148	52.26	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	104.52%
7) Chloroethane-d5	1.569	69	275902	54.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	109.24%
11) 1,1-Dichloroethene-d2	2.110	63	354342	37.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.08%
21) 2-Butanone-d5	3.894	46	462292	111.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	111.57%
24) Chloroform-d	4.354	84	605664	51.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.42%
26) 1,2-Dichloroethane-d4	5.036	65	412897	53.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.96%
32) Benzene-d6	5.055	84	1214587	51.52	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.04%
36) 1,2-Dichloropropane-d6	6.074	67	377295	51.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	103.88%
41) Toluene-d8	7.318	98	1118883	50.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.94%
43) trans-1,3-Dichloroprop...	7.624	79	181221	50.91	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	101.82%
47) 2-Hexanone-d5	8.093	63	339057	105.66	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	105.66%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	513525	52.06	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	104.12%
64) 1,2-Dichlorobenzene-d4	11.630	152	482066	52.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.04%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012821\
Data File : VV020190.D
Acq On : 27 Jan 2021 14:28
Operator : SY/MD
Sample : VV0128WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK89

Quant Time: Jan 28 05:48:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 28 05:47:43 2021
Response via : Initial Calibration

