

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012821\
 Data File : VW020203.D
 Acq On : 27 Jan 2021 19:42
 Operator : SY/MD
 Sample : M1201-07
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jan 28 05:51:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 28 05:47:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	901494	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.859	117	854728	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	432090	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.306	65	317378	50.69	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 101.38%			
7) Chloroethane-d5	1.569	69	259759	53.93	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 107.86%			
11) 1,1-Dichloroethene-d2	2.110	63	395547	43.36	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 86.72%			
21) 2-Butanone-d5	3.894	46	433047	109.60	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 109.60%			
24) Chloroform-d	4.354	84	570240	50.56	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 101.12%			
26) 1,2-Dichloroethane-d4	5.036	65	397097	54.45	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 108.90%			
32) Benzene-d6	5.055	84	1137180	51.41	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 102.82%			
36) 1,2-Dichloropropane-d6	6.074	67	356539	52.31	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 104.62%			
41) Toluene-d8	7.318	98	1023308	49.68	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 99.36%			
43) trans-1,3-Dichloroprop...	7.624	79	162822	48.75	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 97.50%			
47) 2-Hexanone-d5	8.093	63	314038	104.30	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 104.30%			
57) 1,1,2,2-Tetrachloroeth...	10.222	84	476554	51.49	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 102.98%			
64) 1,2-Dichlorobenzene-d4	11.630	152	435141	52.61	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 105.22%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012821\
Data File : VV020203.D
Acq On : 27 Jan 2021 19:42
Operator : SY/MD
Sample : M1201-07
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jan 28 05:51:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 28 05:47:43 2021
Response via : Initial Calibration

