

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW012821\
 Data File : VW020204.D
 Acq On : 27 Jan 2021 20:06
 Operator : SY/MD
 Sample : M1258-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Jan 28 05:51:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jan 28 05:47:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.621	114	821043	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.859	117	789297	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.257	152	416623	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.306	65	267305	46.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.74%
7) Chloroethane-d5	1.570	69	199330	45.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.88%
11) 1,1-Dichloroethene-d2	2.110	63	342687	41.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.50%
21) 2-Butanone-d5	3.888	46	346539	96.30	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.30%
24) Chloroform-d	4.354	84	484607	47.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.36%
26) 1,2-Dichloroethane-d4	5.036	65	321221	48.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.72%
32) Benzene-d6	5.055	84	998866	48.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.80%
36) 1,2-Dichloropropane-d6	6.074	67	304623	48.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.80%
41) Toluene-d8	7.318	98	915110	48.11	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.22%
43) trans-1,3-Dichloroprop...	7.624	79	142841	46.31	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.62%
47) 2-Hexanone-d5	8.090	63	271641	97.70	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.70%
57) 1,1,2,2-Tetrachloroeth...	10.222	84	425382	49.78	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	99.56%
64) 1,2-Dichlorobenzene-d4	11.630	152	422008	52.92	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.84%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV012821\
Data File : VV020204.D
Acq On : 27 Jan 2021 20:06
Operator : SY/MD
Sample : M1258-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Jan 28 05:51:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM012821WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jan 28 05:47:43 2021
Response via : Initial Calibration

