

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012825\
 Data File : VV038560.D
 Acq On : 28 Jan 2025 14:55
 Operator : SY/MD
 Sample : VSTDCCC0.5EC
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5309

Quant Time: Jan 29 09:25:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5138	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5145	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2395	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	2527	0.488	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	98.000%
4) 1,2-Dichloroethane-d4	4.969	65	2190	0.529	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.000%
7) 1,2-Dichloropropane-d6	5.995	67	2590	0.512	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	102.000%
8) Toluene-d8	7.253	98	4814	0.475	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	94.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2280	0.529	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	106.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	3289	0.492	ug/L	97
6) Trichloroethene	5.852	95	2093	0.492	ug/L	99
9) 1,2-Dibromoethane	8.305	107	1531	0.508	ug/L	99
12) 1,2,3-Trichloropropane	10.197	75	1369	0.499	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.374	75	225	0.522	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012825\
Data File : VV038560.D
Acq On : 28 Jan 2025 14:55
Operator : SY/MD
Sample : VSTDCCC0.5EC
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5309

Quant Time: Jan 29 09:25:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

