

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012825\
 Data File : VV038555.D
 Acq On : 28 Jan 2025 11:57
 Operator : SY/MD
 Sample : VV0128WBL02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK219

Quant Time: Jan 29 09:24:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5565	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5542	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2603	0.500	ug/L	0.00

System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	2573	0.458	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	92.000%
4) 1,2-Dichloroethane-d4	4.969	65	2219	0.495	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
7) 1,2-Dichloropropane-d6	5.995	67	2623	0.482	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	96.000%
8) Toluene-d8	7.253	98	4896	0.448	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	90.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2384	0.513	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012825\
Data File : VV038555.D
Acq On : 28 Jan 2025 11:57
Operator : SY/MD
Sample : VV0128WBL02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK219

Quant Time: Jan 29 09:24:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

