

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012925\
 Data File : VV038566.D
 Acq On : 29 Jan 2025 12:24
 Operator : SY/MD
 Sample : MDL07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL07

Quant Time: Jan 30 08:33:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Jan 28 08:06:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	5330	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	5344	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	2550	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.273	65	2582	0.480	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	96.000%
4) 1,2-Dichloroethane-d4	4.969	65	2189	0.510	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
7) 1,2-Dichloropropane-d6	5.995	67	2604	0.496	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	100.000%
8) Toluene-d8	7.253	98	4886	0.464	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	92.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	2282	0.510	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.000%
Target Compounds						
3) Vinyl chloride	1.280	62	201	0.029	ug/L #	36
6) Trichloroethene	5.876	95	99	0.022	ug/L #	92
9) 1,2-Dibromoethane	8.336	107	75	0.024	ug/L #	78
12) 1,2,3-Trichloropropane	10.214	75	81	0.028	ug/L	94
13) 1,2-Dibromo-3-chloropr...	12.391	75	15	0.033	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV012925\
Data File : VV038566.D
Acq On : 29 Jan 2025 12:24
Operator : SY/MD
Sample : MDL07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL07

Quant Time: Jan 30 08:33:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM012725.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Jan 28 08:06:10 2025
Response via : Initial Calibration

