

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV013020\
 Data File : VV014433.D
 Acq On : 30 Jan 2020 14:05
 Operator : SY/MD
 Sample : L1324-21
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GAT75

Quant Time: Jan 31 08:03:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR012920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Jan 31 07:45:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	255223	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	237211	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	108990	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	83302	5.65	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	113.00%
7) Chloroethane-d5	1.59	69	51384	4.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	2.14	63	75512	3.51	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.20%
20) 2-Butanone-d5	3.93	46	115142	38.69	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.38%
24) Chloroform-d	4.40	84	125336	3.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.60%
26) 1,2-Dichloroethane-d4	5.08	65	59833	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
32) Benzene-d6	5.10	84	285772	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
36) 1,2-Dichloropropane-d6	6.11	67	89242	4.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.40%
41) Toluene-d8	7.36	98	261295	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.40%
43) trans-1,3-Dichloropropene-	7.66	79	34050	4.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.00%
46) 2-Hexanone-d5	8.15	63	115548	37.72	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.44%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	58774	4.22	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	87614	4.48	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.60%

Target Compounds

					Ovalue
13) Acetone	2.19	43	7744	4.258 ug/L	69
14) Carbon disulfide	2.32	76	14007	0.267 ug/L	97
25) Chloroform	4.42	83	7289	0.224 ug/L	89
42) Toluene	7.43	91	8428	0.107 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV013020\
Data File : VV014433.D
Acq On : 30 Jan 2020 14:05
Operator : SY/MD
Sample : L1324-21
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GAT75

Quant Time: Jan 31 08:03:14 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR012920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jan 31 07:45:41 2020
Response via : Initial Calibration

