

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013024\
 Data File : VV033905.D
 Acq On : 30 Jan 2024 10:07
 Operator : SY/MD
 Sample : VV0130WBL01
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK204

Quant Time: Jan 30 13:57:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 30 03:43:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	98029	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	88383	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	44605	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	32740	4.312	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.200%
7) Chloroethane-d5	1.536	69	32443	4.879	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.600%
11) 1,1-Dichloroethene-d2	2.060	65	15589	4.258	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.200%
20) 2-Butanone-d5	3.828	46	69280m	54.334	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.660%
24) Chloroform-d	4.253	84	69183	5.015	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.200%
26) 1,2-Dichloroethane-d4	4.947	65	34633	5.372	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.400%
32) Benzene-d6	4.963	84	136284	5.139	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.800%
36) 1,2-Dichloropropane-d6	5.992	67	41159	5.553	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.000%
41) Toluene-d8	7.243	98	118229	4.851	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.000%
43) trans-1,3-Dichloroprop...	7.561	79	13182	4.710	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.200%
46) 2-Hexanone-d5	8.031	63	51925	51.969	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.940%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	23765	5.289	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.800%
66) 1,2-Dichlorobenzene-d4	11.561	152	41988	5.176	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013024\
Data File : VV033905.D
Acq On : 30 Jan 2024 10:07
Operator : SY/MD
Sample : VV0130WBL01
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK204

Quant Time: Jan 30 13:57:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 30 03:43:29 2024
Response via : Initial Calibration

