

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW013024\
 Data File : VW033906.D
 Acq On : 30 Jan 2024 10:49
 Operator : SY/MD
 Sample : VW0130WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK205

Quant Time: Jan 30 14:00:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Jan 30 03:43:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	106303	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	98111	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	50040	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	33237	4.037	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.800%
7) Chloroethane-d5	1.536	69	32370	4.489	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.800%
11) 1,1-Dichloroethene-d2	2.060	65	16048	4.042	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.800%
20) 2-Butanone-d5	3.831	46	58859	42.568	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	85.140%
24) Chloroform-d	4.249	84	66740	4.461	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.200%
26) 1,2-Dichloroethane-d4	4.944	65	31159	4.457	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.200%
32) Benzene-d6	4.960	84	130788	4.443	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	5.989	67	38401	4.667	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.400%
41) Toluene-d8	7.243	98	119018	4.399	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.000%
43) trans-1,3-Dichloroprop...	7.558	79	12733	4.098	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.000%
46) 2-Hexanone-d5	8.030	63	44404	40.035	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.080%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	21772	4.365	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.200%
66) 1,2-Dichlorobenzene-d4	11.561	152	41826	4.596	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013024\
Data File : VV033906.D
Acq On : 30 Jan 2024 10:49
Operator : SY/MD
Sample : VV0130WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK205

Quant Time: Jan 30 14:00:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR012524WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Jan 30 03:43:29 2024
Response via : Initial Calibration

