

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV013025\
 Data File : VV038570.D
 Acq On : 29 Jan 2025 14:45
 Operator : SY/MD
 Sample : VV0130WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK279

Quant Time: Jan 30 08:17:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 30 08:15:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	159673	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.777	117	172713	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.182	152	68477	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.275	65	49847	3.786	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	75.800%	
7) Chloroethane-d5	1.529	69	50209	4.541	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.800%	
11) 1,1-Dichloroethene-d2	2.050	65	25652	4.323	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	86.400%	
20) 2-Butanone-d5	3.835	46	172293	62.212	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	124.420%	
24) Chloroform-d	4.240	84	135891	5.506	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	110.200%	
26) 1,2-Dichloroethane-d4	4.937	65	63475	5.613	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	112.200%	
32) Benzene-d6	4.950	84	226539	4.762	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.200%	
36) 1,2-Dichloropropane-d6	5.979	67	78543	5.432	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	108.600%	
41) Toluene-d8	7.236	98	165143	3.865	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	77.400%	
43) trans-1,3-Dichloroprop...	7.558	79	24894	4.707	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.200%	
46) 2-Hexanone-d5	8.024	63	92649	44.401	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	88.800%	
56) 1,1,2,2-Tetrachloroeth...	10.146	84	67933	5.926	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	118.600%	
66) 1,2-Dichlorobenzene-d4	11.555	152	64586	5.573	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	111.400%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV013025\
 Data File : VV038570.D
 Acq On : 29 Jan 2025 14:45
 Operator : SY/MD
 Sample : VV0130WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK279

Quant Time: Jan 30 08:17:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR011725WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Jan 30 08:15:13 2025
 Response via : Initial Calibration

