

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
 Data File : VV020229.D
 Acq On : 29 Jan 2021 18:46
 Operator : SY/MD
 Sample : L5214-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL-WATER-QT1-2021-02

Manual Integrations
 APPROVED

Quant Time: Jan 30 03:50:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jan 30 03:45:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	12500	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	11880	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	6042	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	7730	0.49	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	98.00%
4) 1,2-Dichloroethane-d4	5.043	65	4763	0.50	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.00%
7) 1,2-Dichloropropane-d6	6.069	67	5648	0.52	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.00%
8) Toluene-d8	7.317	98	14878	0.49	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.00%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	3238	0.50	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	100.00%
Target Compounds						
3) Vinyl chloride	1.310	62	461	0.03	ug/L #	30
6) Trichloroethene	5.926	95	283	0.03	ug/L	90
9) 1,2-Dibromoethane	8.369	107	129	0.02	ug/L #	87
12) 1,2,3-Trichloropropane	10.267	75	132	0.03	ug/L	95
13) 1,2-Dibromo-3-chloropr...	12.427	75	29m	0.03	ug/L	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

MMDadoda


2/5/2021 12:34:18 PM

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
Data File : VV020229.D
Acq On : 29 Jan 2021 18:46
Operator : SY/MD
Sample : L5214-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL-WATER-QT1-2021-02

Manual Integrations
APPROVED

Quant Time: Jan 30 03:50:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jan 30 03:45:01 2021
Response via : Initial Calibration



MMDadoda

2/5/2021 12:34:18 PM