

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
 Data File : VV020224.D
 Acq On : 29 Jan 2021 15:18
 Operator : SY/MD
 Sample : VSTD0.531
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5208

Quant Time: Jan 30 03:38:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jan 30 03:37:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	15208	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	14133	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	7410	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	9148	0.66	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	5578	0.45	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.068	67	6513	0.46	ug/L	0.00
8) Toluene-d8	7.317	98	17990	0.53	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	3713	0.45	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.303	62	8083	0.38	ug/L	100
6) Trichloroethene	5.925	95	5509	0.40	ug/L	100
9) 1,2-Dibromoethane	8.368	107	3148	0.42	ug/L	100
12) 1,2,3-Trichloropropane	10.267	75	2945	0.39	ug/L	100
13) 1,2-Dibromo-3-chloropr...	12.426	75	451	0.41	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
Data File : VV020224.D
Acq On : 29 Jan 2021 15:18
Operator : SY/MD
Sample : VSTD0.531
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5208

Quant Time: Jan 30 03:38:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jan 30 03:37:50 2021
Response via : Initial Calibration

