

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
 Data File : VV020225.D
 Acq On : 29 Jan 2021 15:41
 Operator : SY/MD
 Sample : VSTD00132
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001209

Quant Time: Jan 30 03:38:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Jan 30 03:37:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	14849	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	14335	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	7366	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	18632	1.37	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	11483	0.96	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.069	67	13600	0.95	ug/L	0.00
8) Toluene-d8	7.317	98	37498	1.10	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	8124	0.97	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	17286	0.83	ug/L	95
6) Trichloroethene	5.925	95	11783	0.84	ug/L	100
9) 1,2-Dibromoethane	8.353	107	6835	0.91	ug/L #	94
12) 1,2,3-Trichloropropane	10.267	75	6477	0.85	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.427	75	1018	0.93	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV013121\
Data File : VV020225.D
Acq On : 29 Jan 2021 15:41
Operator : SY/MD
Sample : VSTD00132
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001209

Quant Time: Jan 30 03:38:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM013121.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Jan 30 03:37:50 2021
Response via : Initial Calibration

